

## PRESTO W50

# Heating a 100 liters reactor from +20 °C to +100 °C

### Objective

This case study tests the heating power of PRESTO W50 with a 100 liters glass reactor. The PRESTO W50 is connected to the reactor via two 2 m metal tubings. The PRESTO W50 is programmed to heat up from +20 °C to +100 °C.

### Environment

Room temperature +20 °C  
 Humidity 45%  
 Voltage 400 V / 50 Hz

### Test Conditions

JULABO unit	PRESTO W50
Cooling power	+20 °C 7.5 kW 0 °C 6.5 kW -20 °C 3.0 kW
Heating capacity	6 kW
Band limit	without
Flow pressure	0.5 bar
Bath fluid	Thermal HL60
Reactor	100 liters glass reactor (Büchiglas) filled with 80 l Thermal HL60
Jacket volume	30 l
Control	External (ICC)

### Control Parameters


Xp	0.2 K
Tn	695 s
Tv	85 s
Xpu	15 K



## Test Results

The PRESTO W50 heating process from +20 °C to +100°C in 1 h 55 min without overshoot.

Measured with *EasyTEMP Professional*



- Setpoint
- Temperature in reactor's interior
- Temperature in reactor's jacket

### Tip

Use our tube adapters and your tubing will no longer kink.



### Tip

You can also use the robust Pt100 with PTFE coating.

